


SERIES 1300 Oxygen Deficiency Monitor¹

Example Installation Configuration

Electronics with Audible and Visual Alarms Located Outside with
Optional Horn & Strobe and Optional Remote Sensor Enclosure Inside

*Note: For Lighter Gases Such as He, Consideration
Should be Given for Installing the Remote Sensor at a
Higher Elevation.*



For Installations Requiring Conduit,
Please Contact Factory

*Note: Horn and Strobe Can be Installed at a
Different Location than the Remote Sensor*

¹ Disclaimer: The above installation is for example purposes only. It does not represent a specific recommendation by Alpha Omega Instruments. The exact placement of the monitor(s) and sensor(s), as well as the number required, is to be determined by the end user based on the physical layout of the proposed installation.