


SERIES 1300 Oxygen Deficiency Monitor¹

Example Installation Configuration

Oxygen Monitor and Optional Remote Sensor Inside Area of Concern for Inert Gases that are Heavier than Air with Optional Horn and Strobe at Entrance



Note: For Inert Gases that are Heavier than Air Such as CO₂ and SF₆, Consideration Should be Given for Installing the Remote Sensor at a Lower Elevation.

For Installations Requiring Conduit,
Please Contact Factory

¹ Disclaimer: The above installation is for example purposes only. It does not represent a specific recommendation by Alpha Omega Instruments. The exact placement of the monitor(s) and sensor(s), as well as the number required, is to be determined by the end user based on the physical layout of the proposed installation.